
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002404
13:00 20/12/2024
AIR DRYING ONLY
60 min. MAXIMUM
CYCLE START 0:00
AIR DRYING 0:08
CYCLE ABORTED 53:34
CASSETTE REMOVED
END TIME 53:34

Digital Signature #
68E6221804563694
