

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2720

Name: RITTER 1  
Sterilizer ID: V2178198  
12/22/2022 02:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
2:00 187.6 F 2.7  
4:00 191.6 F 7.8  
6:00 212.0 F 13.5  
8:00 226.1 F 8.2  
10:00 239.1 F 12.6  
12:00 239.9 F 11.3  
14:00 238.6 F 9.7  
16:00 252.7 F 16.8  
18:00 264.8 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.5 F 29.0  
2:00 271.6 F 28.5  
3:00 272.5 F 29.0  
4:00 271.6 F 28.3  
  
Min 271.5 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.9 F 0.0  
5:00 223.3 F 0.0  
10:00 228.2 F 0.0  
15:00 230.7 F 0.0  
20:00 227.8 F 0.0  
25:00 228.7 F 0.0  
30:00 229.6 F 0.0  
  
Drying Complete

FILLING: 1:59  
HEATING: 19:48  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:00  
CYCLE COMPLETE

12/22/2022 03:03 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_