

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3027

Name: RITTER 1  
Sterilizer ID: V2178198  
03/22/2023 01:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.4 F 0.0  
2:00 190.1 F 4.0  
4:00 202.7 F 9.0  
6:00 221.5 F 7.6  
8:00 233.6 F 10.1  
10:00 235.8 F 7.6  
12:00 243.7 F 12.2  
14:00 245.7 F 13.0  
16:00 259.0 F 20.2  
18:00 270.6 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 28.9  
2:00 271.4 F 28.2  
3:00 272.6 F 29.0  
4:00 271.8 F 28.4  
  
Min 271.4 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 217.9 F 0.0  
10:00 222.7 F 0.0  
15:00 223.7 F 0.0  
20:00 225.0 F 0.0  
25:00 224.7 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 18:47  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:35  
CYCLE COMPLETE

03/22/2023 02:22 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_