
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000880
15:28 21/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:51
STERILIZING 5:04
136.0 C 225kPa 5:04
136.4 C 229kPa 5:34
136.2 C 230kPa 6:04
136.2 C 231kPa 6:34
136.2 C 230kPa 7:04
136.2 C 231kPa 7:34
136.2 C 231kPa 8:04
136.4 C 233kPa 8:34
136.4 C 233kPa 8:34
VENTING 8:34
AIR DRYING 9:28
CYCLE COMPLETE 18:16

Digital Signature #
1D555F15DD7DFA8A
