

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3119

Name: RITTER 1  
Sterilizer ID: V2178198  
04/18/2023 07:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.8 F 0.0  
2:00 80.8 F 0.0  
4:00 88.6 F 0.0  
6:00 117.8 F 0.0  
8:00 147.4 F 0.0  
10:00 184.4 F 0.0  
12:00 207.6 F 3.2  
14:00 215.2 F 6.2  
16:00 223.6 F 9.2  
18:00 232.1 F 12.9  
20:00 232.3 F 8.6  
22:00 243.6 F 13.5  
24:00 241.4 F 11.3  
26:00 237.8 F 9.6  
28:00 251.5 F 16.0  
30:00 262.7 F 22.4  
32:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.4 F 28.7  
2:00 272.1 F 28.8  
3:00 271.9 F 28.4  
4:00 272.5 F 29.1  
  
Min 271.4 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
5:00 208.9 F 0.0  
10:00 229.7 F 0.0  
15:00 231.7 F 0.0  
20:00 224.4 F 0.0  
25:00 228.4 F 0.0  
30:00 220.4 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 32:20  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:10:24  
CYCLE COMPLETE

04/18/2023 08:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_