
Midmark M11 - v1.0.0
Total Cycles: 5432

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2025 03:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.1 F 0.0
2:00 120.9 F 1.6
4:00 145.2 F 4.8
6:00 172.2 F 8.8
8:00 194.3 F 13.4
10:00 209.4 F 4.9
12:00 223.6 F 9.3
14:00 230.8 F 9.7
16:00 237.2 F 10.6
18:00 234.3 F 8.2
20:00 249.0 F 15.1
22:00 261.1 F 21.9
24:00 272.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 28.8
2:00 272.7 F 29.3
3:00 271.8 F 28.9
4:00 272.4 F 29.0

Min 271.5 F 28.6
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.7 F 0.0
5:00 188.6 F 0.0
10:00 193.4 F 0.0
15:00 214.2 F 0.0
20:00 221.5 F 0.0
25:00 231.7 F 0.0
30:00 238.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:22
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:02:47
CYCLE COMPLETE

03/21/2025 04:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____