
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000335
10:26 08/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:56
STERILIZING 3:04
136.0 C 227kPa 3:04
136.4 C 232kPa 3:34
136.3 C 231kPa 4:04
136.4 C 233kPa 4:34
136.4 C 233kPa 5:04
136.2 C 232kPa 5:34
136.1 C 229kPa 6:04
136.3 C 231kPa 6:34
136.3 C 231kPa 6:34
VENTING 6:34
AIR DRYING 7:30
CYCLE COMPLETE 22:30

Digital Signature #
0BF98770F800B5DD
