
Midmark M11 - v1.0.0
Total Cycles: 4301

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2024 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.8 F 0.0
2:00 150.8 F 1.6
4:00 169.1 F 4.5
6:00 179.1 F 8.1
8:00 196.3 F 12.0
10:00 209.9 F 3.8
12:00 220.2 F 7.5
14:00 230.9 F 11.5
16:00 231.7 F 7.2
18:00 242.0 F 12.4
20:00 240.1 F 10.5
22:00 251.9 F 16.2
24:00 262.3 F 22.2
26:00 272.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 271.9 F 28.7
2:00 272.6 F 28.9
3:00 271.8 F 28.4
4:00 271.8 F 28.6

Min 271.4 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 215.0 F 0.0
10:00 230.7 F 0.0
15:00 224.2 F 0.0
20:00 216.1 F 0.0
25:00 211.8 F 0.0
30:00 208.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:27
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:05:07
CYCLE COMPLETE

02/14/2024 12:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____