

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4470

Name: RITTER2  
Sterilizer ID: V2313228  
03/28/2025 10:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.5 F 0.0  
2:00 131.4 F 1.1  
4:00 135.6 F 3.3  
6:00 149.5 F 6.6  
8:00 170.0 F 10.8  
10:00 211.9 F 4.7  
12:00 218.8 F 7.9  
14:00 225.6 F 12.2  
16:00 232.5 F 8.8  
18:00 240.9 F 11.5  
20:00 244.1 F 12.5  
22:00 254.8 F 18.4  
24:00 264.7 F 24.8  
26:00 270.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 30.2  
1:00 272.1 F 30.5  
2:00 272.0 F 30.2  
3:00 271.8 F 30.1  
4:00 271.9 F 30.4

Min 271.8 F 30.1  
Max 272.3 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.6 F 0.0  
5:00 190.9 F 0.0  
10:00 202.3 F 0.0  
15:00 199.6 F 0.0  
20:00 196.8 F 0.0  
25:00 196.6 F 0.0  
30:00 200.2 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 26:42  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:05:27  
CYCLE COMPLETE

03/28/2025 11:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_