

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000113  
11:22 03/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:23  
PRESSURIZING 3:25  
STERILIZING 4:56  
136.0 C 221kPa 4:56  
136.2 C 222kPa 5:26  
136.4 C 225kPa 5:56  
136.4 C 226kPa 6:26  
136.3 C 226kPa 6:56  
136.3 C 226kPa 7:26  
136.3 C 226kPa 7:56  
136.4 C 226kPa 8:26  
136.4 C 226kPa 8:26  
VENTING 8:26  
AIR DRYING 9:29  
CYCLE COMPLETE 24:29

-----  
Digital Signature #  
7F45ED4103D91088  
-----