

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002160  
9:36 22/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:29  
PRESSURIZING 3:31  
STERILIZING 4:48  
136.0 C 223kPa 4:48  
136.4 C 226kPa 5:18  
136.4 C 225kPa 5:48  
136.4 C 225kPa 6:18  
136.4 C 226kPa 6:48  
136.4 C 226kPa 7:18  
136.4 C 226kPa 7:48  
136.3 C 224kPa 8:18  
136.3 C 224kPa 8:18  
VENTING 8:18  
AIR DRYING 9:11  
CYCLE COMPLETE 9:51  
-----

Digital Signature #  
E2254AAFD295A53D  
-----