

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1737

Name: RITTER2  
Sterilizer ID: V2313228  
11/16/2022 05:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.1 F 0.0  
2:00 158.8 F 0.2  
4:00 190.7 F 0.8  
6:00 209.6 F 5.3  
8:00 229.6 F 10.4  
10:00 231.9 F 7.3  
12:00 241.7 F 13.1  
14:00 241.7 F 11.9  
16:00 241.7 F 11.5  
18:00 255.6 F 18.7  
20:00 267.6 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.5 F 29.3  
2:00 272.0 F 29.4  
3:00 272.3 F 29.1  
4:00 272.0 F 29.5

Min 271.5 F 28.8  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 213.7 F 0.0  
10:00 230.1 F 0.0  
15:00 225.9 F 0.0  
20:00 224.1 F 0.0  
25:00 227.8 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:48  
CYCLE COMPLETE

11/16/2022 06:07 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_