
Midmark M11 - v1.0.5
Total Cycles: 540

Name: RITTER2
Sterilizer ID: V2313228
11/11/2021 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 175.9 F 2.8
4:00 195.7 F 7.8
6:00 217.7 F 11.2
8:00 226.7 F 9.4
10:00 234.4 F 9.2
12:00 244.8 F 13.5
14:00 244.8 F 12.8
16:00 257.7 F 19.8
18:00 269.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.0
2:00 272.0 F 29.5
3:00 271.6 F 28.7
4:00 272.4 F 29.5

Min 271.2 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 230.8 F 0.0
10:00 193.2 F 0.0
15:00 178.4 F 0.0
20:00 171.8 F 0.0
25:00 169.8 F 0.0
30:00 184.1 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 19:00
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:03
CYCLE COMPLETE

11/11/2021 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____