
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000280
15:12 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:24
STERILIZING 4:50
136.0 C 229kPa 4:50
136.4 C 233kPa 5:20
136.2 C 232kPa 5:50
136.3 C 233kPa 6:20
136.2 C 232kPa 6:50
136.4 C 233kPa 7:20
136.3 C 233kPa 7:50
136.4 C 232kPa 8:20
136.4 C 232kPa 8:20
VENTING 8:20
AIR DRYING 9:23
CYCLE COMPLETE 22:36

Digital Signature #
4A899F8D7C4E2F5F
