
Midmark M11 - v1.0.5
Total Cycles: 1655

Name: RITTER2
Sterilizer ID: V2313228
10/25/2022 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.7 F 0.0
2:00 192.6 F 3.8
4:00 199.2 F 9.5
6:00 222.3 F 7.2
8:00 234.8 F 11.4
10:00 235.3 F 9.3
12:00 238.7 F 9.4
14:00 248.8 F 15.3
16:00 261.3 F 22.3
18:00 272.6 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.5 F 29.6
2:00 271.8 F 29.4
3:00 272.4 F 29.5
4:00 271.7 F 29.3

Min 271.6 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 212.8 F 0.0
10:00 221.3 F 0.0
15:00 226.1 F 0.0
20:00 227.6 F 0.0
25:00 229.6 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 3:21
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

10/25/2022 02:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____