
Midmark M11 - v1.0.5
Total Cycles: 1311

Name: RITTER2
Sterilizer ID: V2313228
07/18/2022 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.7 F 0.0
2:00 185.9 F 0.2
4:00 190.0 F 2.9
6:00 206.6 F 7.7
8:00 224.3 F 12.9
10:00 230.7 F 9.1
12:00 237.8 F 10.4
14:00 244.3 F 13.4
16:00 244.6 F 12.9
18:00 257.6 F 19.9
20:00 269.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.7
2:00 271.9 F 28.9
3:00 272.2 F 29.6
4:00 272.5 F 29.5

Min 271.4 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 224.0 F 0.0
10:00 238.1 F 0.0
15:00 233.7 F 0.0
20:00 225.0 F 0.0
25:00 222.7 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

07/18/2022 11:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____