
Midmark M11 - v1.0.5
Total Cycles: 1764

Name: RITTER2
Sterilizer ID: V2313228
11/22/2022 04:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 171.6 F 0.3
4:00 177.2 F 2.5
6:00 200.7 F 6.7
8:00 216.6 F 11.1
10:00 224.0 F 6.6
12:00 236.7 F 12.1
14:00 237.3 F 10.2
16:00 237.3 F 8.9
18:00 250.5 F 16.0
20:00 262.1 F 22.7
22:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 271.7 F 29.4
2:00 272.7 F 29.7
3:00 272.1 F 29.2
4:00 271.3 F 28.9

Min 271.3 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 211.4 F 0.0
10:00 216.6 F 0.0
15:00 221.0 F 0.0
20:00 217.5 F 0.0
25:00 215.5 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

11/22/2022 05:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____