
Midmark M11 - v1.0.0
Total Cycles: 3384

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2023 07:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 77.6 F 0.0
4:00 92.8 F 0.0
6:00 111.0 F 0.0
8:00 138.3 F 1.4
10:00 168.5 F 4.2
12:00 193.9 F 7.0
14:00 201.3 F 11.2
16:00 213.8 F 4.4
18:00 222.4 F 7.7
20:00 232.1 F 11.7
22:00 230.3 F 7.7
24:00 242.9 F 13.0
26:00 240.3 F 10.9
28:00 252.2 F 16.8
30:00 262.5 F 22.9
32:00 272.0 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.6
2:00 272.1 F 29.0
3:00 271.4 F 28.8
4:00 272.5 F 29.6

Min 271.2 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 204.7 F 0.0
10:00 209.0 F 0.0
15:00 195.0 F 0.0
20:00 200.2 F 0.0
25:00 201.8 F 0.0
30:00 204.2 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 32:33
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:10:36
CYCLE COMPLETE

06/14/2023 08:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____