
Midmark M11 - v1.0.5
Total Cycles: 3633

Name: RITTER2
Sterilizer ID: V2313228
05/22/2024 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 70.2 F 0.0
4:00 76.7 F 0.3
6:00 102.9 F 1.9
8:00 136.8 F 4.6
10:00 164.6 F 7.5
12:00 195.4 F 10.8
14:00 203.7 F 5.2
16:00 215.8 F 7.1
18:00 226.0 F 10.9
20:00 227.4 F 7.1
22:00 238.5 F 11.8
24:00 237.3 F 9.8
26:00 249.7 F 15.8
28:00 260.3 F 21.8
30:00 270.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.5 F 30.1
2:00 271.8 F 29.3
3:00 271.7 F 29.7
4:00 272.5 F 29.9

Min 271.1 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 214.0 F 0.0
10:00 205.4 F 0.0
15:00 204.2 F 0.0
20:00 203.6 F 0.0
25:00 205.7 F 0.0
30:00 206.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:60
CYCLE COMPLETE

05/22/2024 09:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____