

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001525  
10:51 21/04/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:35  
STERILIZING 3:37  
136.5 C 224kPa 3:37  
136.3 C 223kPa 4:07  
136.2 C 223kPa 4:37  
136.2 C 222kPa 5:07  
136.3 C 222kPa 5:37  
136.2 C 222kPa 6:07  
136.4 C 222kPa 6:37  
136.3 C 221kPa 7:07  
136.3 C 221kPa 7:07  
VENTING 7:07  
CYCLE ABORTED 7:54  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 7:54  
-----

Digital Signature #  
6A8875282C9AD18E  
-----