
Midmark M11 - v1.0.5
Total Cycles: 3828

Name: RITTER2
Sterilizer ID: V2313228
08/01/2024 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.6 F 0.0
2:00 92.5 F 0.0
4:00 117.4 F 0.1
6:00 143.7 F 1.3
8:00 171.2 F 4.0
10:00 192.0 F 6.9
12:00 201.0 F 10.1
14:00 210.8 F 9.3
16:00 221.0 F 7.4
18:00 231.9 F 11.6
20:00 232.0 F 8.2
22:00 243.7 F 13.5
24:00 242.9 F 12.2
26:00 254.3 F 18.1
28:00 264.6 F 24.3
30:00 271.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.6 F 29.5
2:00 271.9 F 29.4
3:00 272.4 F 29.3
4:00 271.9 F 29.4

Min 271.6 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 207.9 F 0.0
10:00 220.8 F 0.0
15:00 220.6 F 0.0
20:00 218.9 F 0.0
25:00 216.8 F 0.0
30:00 215.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:03
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:08:38
CYCLE COMPLETE

08/01/2024 02:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____