
Midmark M11 - v1.0.0
Total Cycles: 2118

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2022 08:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.2 F 0.0
2:00 116.7 F 0.0
4:00 149.8 F 1.1
6:00 168.8 F 4.0
8:00 189.5 F 7.5
10:00 206.2 F 11.6
12:00 216.5 F 4.4
14:00 228.2 F 8.8
16:00 239.1 F 13.4
18:00 237.5 F 10.0
20:00 237.6 F 8.8
22:00 246.6 F 13.5
24:00 257.7 F 19.5
26:00 267.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.3 F 28.6
2:00 272.5 F 29.1
3:00 271.5 F 28.3
4:00 272.3 F 28.8

Min 271.4 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.6 F 0.0
5:00 233.5 F 0.0
10:00 230.5 F 0.0
15:00 223.4 F 0.0
20:00 220.3 F 0.0
25:00 223.2 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 27:21
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:05:11
CYCLE COMPLETE

07/13/2022 09:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____