

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2872

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2023 02:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.2 F 0.0  
2:00 159.7 F 2.9  
4:00 179.2 F 6.8  
6:00 199.8 F 11.5  
8:00 217.8 F 6.3  
10:00 234.5 F 11.5  
12:00 232.5 F 8.4  
14:00 241.4 F 11.6  
16:00 244.7 F 13.0  
18:00 256.6 F 19.3  
20:00 267.3 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 271.2 F 29.3  
2:00 271.1 F 29.3  
3:00 271.2 F 29.2  
4:00 271.7 F 29.3  
  
Min 271.1 F 29.1  
Max 272.3 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
5:00 247.3 F 0.0  
10:00 226.4 F 0.0  
15:00 228.0 F 0.0  
20:00 203.3 F 0.0  
25:00 182.8 F 0.0  
30:00 210.0 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 21:21  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:00:32  
CYCLE COMPLETE

09/12/2023 03:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_