
Midmark M11 - v1.0.5
Total Cycles: 4017

Name: RITTER2
Sterilizer ID: V2313228
10/10/2024 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.5 F 0.0
2:00 154.8 F 1.4
4:00 162.2 F 4.1
6:00 173.7 F 8.2
8:00 194.5 F 12.5
10:00 213.2 F 6.1
12:00 225.4 F 10.8
14:00 228.3 F 7.2
16:00 239.9 F 12.4
18:00 239.0 F 10.7
20:00 251.5 F 16.9
22:00 262.6 F 23.5
24:00 272.6 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.2
1:00 272.1 F 30.1
2:00 273.2 F 30.3
3:00 271.8 F 29.7
4:00 273.0 F 30.5

Min 270.9 F 29.1
Max 273.2 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 192.0 F 0.0
10:00 221.3 F 0.0
15:00 226.1 F 0.0
20:00 227.4 F 0.0
25:00 230.0 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:21
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:01
CYCLE COMPLETE

10/10/2024 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____