

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001273  
10:06 27/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CYCLE ABORTED 0:13  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 0:13  
-----

Digital Signature #  
E7F0D4AB784B8D00  
-----