
Midmark M11 - v1.0.0
Total Cycles: 1285

Name: RITTER 1
Sterilizer ID: V2178198
10/21/2021 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.9 F 0.0
2:00 190.2 F 3.3
4:00 208.0 F 8.9
6:00 222.6 F 7.9
8:00 233.5 F 10.1
10:00 234.2 F 8.2
12:00 243.8 F 11.9
14:00 246.1 F 13.2
16:00 259.2 F 20.3
18:00 270.8 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.5 F 28.7
2:00 272.0 F 28.6
3:00 272.5 F 28.7
4:00 271.7 F 28.4

Min 271.5 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 220.7 F 0.0
10:00 201.0 F 0.0
15:00 196.0 F 0.0
20:00 193.0 F 0.0
25:00 192.3 F 0.0
30:00 192.3 F 0.0
Drying Complete

FILLING: 2:49
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

10/21/2021 11:22 AM
ACCEPTED: X
REJECTED:
Operator: U02