
Midmark M11 - v1.0.5
Total Cycles: 4252

Name: RITTER2
Sterilizer ID: V2313228
01/09/2025 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.8 F 0.0
2:00 95.0 F 0.0
4:00 111.3 F 0.1
6:00 139.8 F 0.5
8:00 162.7 F 1.3
10:00 186.6 F 4.3
12:00 204.6 F 7.8
14:00 217.8 F 11.6
16:00 223.9 F 6.4
18:00 233.9 F 11.3
20:00 233.9 F 8.8
22:00 241.9 F 11.1
24:00 246.0 F 13.6
26:00 257.5 F 19.8
28:00 267.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.3
2:00 271.8 F 29.6
3:00 272.0 F 29.0
4:00 272.5 F 29.7

Min 271.2 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 240.3 F 0.0
10:00 221.0 F 0.0
15:00 223.5 F 0.0
20:00 223.8 F 0.0
25:00 228.8 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:19
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:58
CYCLE COMPLETE

01/09/2025 04:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____