
Midmark M11 - v1.0.5
Total Cycles: 840

Name: RITTER2
Sterilizer ID: V2313228
02/24/2022 03:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.7 F 0.0
2:00 169.3 F 2.9
4:00 190.4 F 7.6
6:00 210.8 F 13.3
8:00 227.4 F 8.4
10:00 238.7 F 11.4
12:00 241.9 F 12.1
14:00 241.4 F 11.2
16:00 255.1 F 18.3
18:00 266.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.1 F 28.7
2:00 272.3 F 29.0
3:00 272.2 F 29.4
4:00 271.4 F 28.7

Min 271.4 F 28.6
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.0 F 0.0
5:00 228.5 F 0.0
10:00 216.8 F 0.0
15:00 228.9 F 0.0
20:00 213.4 F 0.0
25:00 221.6 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

02/24/2022 05:08 PM
ACCEPTED: X
REJECTED:
Operator: U02