
Midmark M11 - v1.0.0
Total Cycles: 4633

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2024 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.8 F 0.0
2:00 166.3 F 3.2
4:00 186.1 F 7.0
6:00 205.3 F 11.3
8:00 222.4 F 4.8
10:00 226.1 F 9.4
12:00 236.6 F 10.7
14:00 240.7 F 11.0
16:00 236.8 F 8.9
18:00 250.1 F 15.1
20:00 260.8 F 21.2
22:00 270.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.6 F 29.0
2:00 272.2 F 29.1
3:00 271.5 F 28.7
4:00 272.0 F 28.7

Min 270.9 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 227.4 F 0.0
10:00 204.5 F 0.0
15:00 203.7 F 0.0
20:00 237.8 F 0.0
25:00 212.5 F 0.0
30:00 236.5 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 22:43
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:02:13
CYCLE COMPLETE

06/12/2024 04:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____