
Midmark M11 - v1.0.5
Total Cycles: 4093

Name: RITTER2
Sterilizer ID: V2313228
11/06/2024 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.7 F 0.0
2:00 87.2 F 0.0
4:00 109.6 F 0.0
6:00 144.9 F 0.1
8:00 169.9 F 0.2
10:00 188.8 F 2.0
12:00 206.5 F 4.6
14:00 214.1 F 7.8
16:00 223.2 F 11.4
18:00 225.9 F 6.4
20:00 235.8 F 11.2
22:00 233.6 F 8.5
24:00 243.1 F 12.0
26:00 244.9 F 13.1
28:00 255.9 F 18.9
30:00 265.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.7
2:00 272.1 F 29.1
3:00 271.4 F 29.1
4:00 272.5 F 29.7

Min 271.3 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 222.6 F 0.0
10:00 204.8 F 0.0
15:00 203.3 F 0.0
20:00 200.8 F 0.0
25:00 205.5 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:53
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:10:39
CYCLE COMPLETE

11/06/2024 04:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____