

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3586

Name: RITTER2  
Sterilizer ID: V2313228  
05/03/2024 08:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 71.1 F 0.0  
4:00 76.4 F 0.1  
6:00 102.5 F 1.6  
8:00 136.3 F 4.2  
10:00 167.6 F 7.0  
12:00 204.3 F 10.9  
14:00 217.0 F 4.7  
16:00 222.6 F 7.7  
18:00 233.1 F 11.7  
20:00 230.4 F 7.3  
22:00 240.3 F 12.3  
24:00 237.9 F 9.9  
26:00 250.1 F 15.8  
28:00 261.0 F 21.9  
30:00 270.8 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.1 F 29.1  
2:00 272.5 F 29.6  
3:00 272.1 F 29.7  
4:00 271.4 F 29.0  
  
Min 271.3 F 29.0  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 225.0 F 0.0  
10:00 184.0 F 0.0  
15:00 182.5 F 0.0  
20:00 176.9 F 0.0  
25:00 174.4 F 0.0  
30:00 172.4 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 31:29  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:09:35  
CYCLE COMPLETE

05/03/2024 09:29 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_