

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000609  
10:51 27/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:19  
STERILIZING 4:37  
136.0 C 221kPa 4:37  
136.2 C 232kPa 5:07  
136.4 C 233kPa 5:37  
136.4 C 232kPa 6:07  
136.2 C 231kPa 6:37  
136.0 C 231kPa 7:07  
136.4 C 233kPa 7:37  
136.3 C 231kPa 8:07  
136.3 C 231kPa 8:07  
VENTING 8:07  
AIR DRYING 9:04  
CYCLE COMPLETE 24:05  
-----

Digital Signature #  
1E13003C3D496F6F  
-----