
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001355
14:26 22/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:57
PRESSURIZING 3:59
STERILIZING 5:10
136.0 C 220kPa 5:10
136.1 C 221kPa 5:40
136.4 C 224kPa 6:10
136.2 C 224kPa 6:40
136.0 C 222kPa 7:10
136.3 C 224kPa 7:40
136.4 C 224kPa 8:10
136.4 C 225kPa 8:40
136.4 C 225kPa 8:40
VENTING 8:40
AIR DRYING 9:34
CYCLE COMPLETE 24:34

Digital Signature #
30D28D8BC957FBD9
