
Midmark M11 - v1.0.0
Total Cycles: 4175

Name: RITTER 1
Sterilizer ID: V2178198
01/03/2024 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.7 F 0.0
2:00 133.9 F 1.6
4:00 145.5 F 4.5
6:00 189.7 F 7.8
8:00 195.4 F 12.1
10:00 214.5 F 4.9
12:00 220.9 F 9.1
14:00 232.3 F 13.5
16:00 235.6 F 9.8
18:00 238.0 F 9.4
20:00 245.5 F 13.3
22:00 256.6 F 19.1
24:00 266.8 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.4 F 29.2
2:00 271.2 F 28.5
3:00 272.6 F 29.7
4:00 271.9 F 29.0

Min 271.0 F 28.5
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 230.8 F 0.0
10:00 230.6 F 0.0
15:00 227.3 F 0.0
20:00 226.5 F 0.0
25:00 226.1 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 25:37
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:37
CYCLE COMPLETE

01/03/2024 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____