

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001991  
9:52 09/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:01  
PRESSURIZING 3:04  
STERILIZING 4:01  
136.0 C 220kPa 4:01  
136.4 C 227kPa 4:31  
136.4 C 225kPa 5:01  
136.4 C 226kPa 5:31  
136.4 C 225kPa 6:01  
136.4 C 222kPa 6:31  
136.4 C 226kPa 7:01  
136.4 C 225kPa 7:31  
136.4 C 225kPa 7:31  
VENTING 7:31  
AIR DRYING 8:24  
CYCLE COMPLETE 8:29

-----  
Digital Signature #  
EB0B18E471E152A1  
-----