
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006888
11:09 01/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:58
STERILIZING 3:00
136.5 C 222kPa 3:00
136.4 C 224kPa 3:30
136.4 C 223kPa 4:00
136.4 C 222kPa 4:30
136.4 C 222kPa 5:00
136.4 C 223kPa 5:30
136.4 C 224kPa 6:00
136.3 C 218kPa 6:30
136.3 C 218kPa 6:30
VENTING 6:30
AIR DRYING 7:23
CYCLE COMPLETE 37:23

Digital Signature #
CB232590595C1010
