

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1132

Name: RITTER2  
Sterilizer ID: V2313228  
05/27/2022 01:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
2:00 190.8 F 3.8  
4:00 201.2 F 8.8  
6:00 223.0 F 8.9  
8:00 234.4 F 10.3  
10:00 235.1 F 8.3  
12:00 242.3 F 11.4  
14:00 246.2 F 13.8  
16:00 258.6 F 20.6  
18:00 269.8 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 271.7 F 29.4  
2:00 273.1 F 30.1  
3:00 272.5 F 29.5  
4:00 271.8 F 29.5  
  
Min 270.9 F 28.7  
Max 273.3 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.6 F 0.0  
5:00 235.7 F 0.0  
10:00 228.0 F 0.0  
15:00 228.3 F 0.0  
20:00 226.6 F 0.0  
25:00 228.3 F 0.0  
30:00 219.9 F 0.0  
Drying Complete

FILLING: 2:19  
HEATING: 18:47  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:20  
CYCLE COMPLETE

05/27/2022 02:54 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_