
Midmark M11 - v1.0.5
Total Cycles: 2177

Name: RITTER2
Sterilizer ID: V2313228
03/27/2023 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 190.6 F 3.4
4:00 206.2 F 9.0
6:00 222.7 F 7.8
8:00 233.8 F 11.1
10:00 235.2 F 9.2
12:00 238.1 F 9.4
14:00 250.2 F 15.7
16:00 262.7 F 23.0
18:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.8 F 29.5
2:00 271.8 F 28.8
3:00 272.1 F 29.5
4:00 272.4 F 29.4

Min 271.3 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 211.3 F 0.0
10:00 219.3 F 0.0
15:00 229.2 F 0.0
20:00 225.8 F 0.0
25:00 221.1 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 18:12
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

03/27/2023 11:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____