

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4222

Name: RITTER2  
Sterilizer ID: V2313228  
12/17/2024 02:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 97.4 F 0.0  
2:00 101.1 F 0.1  
4:00 124.4 F 1.5  
6:00 150.6 F 4.3  
8:00 176.8 F 7.6  
10:00 194.0 F 11.1  
12:00 213.5 F 5.4  
14:00 220.2 F 9.3  
16:00 229.2 F 11.1  
18:00 234.6 F 10.2  
20:00 235.6 F 8.9  
22:00 246.3 F 13.8  
24:00 257.1 F 19.8  
26:00 267.7 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.7 F 30.0  
2:00 272.0 F 29.4  
3:00 271.0 F 29.3  
4:00 272.5 F 30.3

Min 271.0 F 29.1  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
5:00 231.1 F 0.0  
10:00 221.7 F 0.0  
15:00 234.5 F 0.0  
20:00 183.8 F 0.0  
25:00 179.3 F 0.0  
30:00 181.2 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:31  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:06:14  
CYCLE COMPLETE

12/17/2024 03:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_