
Midmark M11 - v1.0.5
Total Cycles: 3138

Name: RITTER2
Sterilizer ID: V2313228
11/17/2023 11:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 167.5 F 1.8
4:00 186.2 F 6.0
6:00 207.0 F 10.4
8:00 220.9 F 6.8
10:00 228.5 F 10.3
12:00 232.2 F 8.3
14:00 243.0 F 13.0
16:00 243.4 F 12.9
18:00 256.1 F 19.4
20:00 267.4 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.4 F 29.7
2:00 271.3 F 29.4
3:00 272.5 F 29.8
4:00 271.2 F 29.1

Min 271.1 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 212.3 F 0.0
10:00 213.1 F 0.0
15:00 210.0 F 0.0
20:00 211.0 F 0.0
25:00 211.6 F 0.0
30:00 212.3 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:03
CYCLE COMPLETE

11/17/2023 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____