

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2173

Name: RITTER2  
Sterilizer ID: V2313228  
03/24/2023 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.3 F 0.0  
2:00 150.6 F 2.3  
4:00 157.9 F 5.1  
6:00 182.1 F 9.4  
8:00 211.4 F 8.1  
10:00 221.8 F 8.4  
12:00 234.8 F 12.4  
14:00 237.4 F 11.1  
16:00 236.6 F 9.5  
18:00 250.4 F 16.2  
20:00 262.3 F 23.1  
22:00 272.8 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.6 F 29.9  
2:00 271.4 F 29.1  
3:00 272.4 F 30.1  
4:00 271.9 F 29.4

Min 271.2 F 29.1  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.5 F 0.0  
5:00 229.0 F 0.0  
10:00 234.3 F 0.0  
15:00 233.7 F 0.0  
20:00 221.7 F 0.0  
25:00 230.0 F 0.0  
30:00 215.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 22:20  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:28  
CYCLE COMPLETE

03/24/2023 10:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_