
Midmark M11 - v1.0.5
Total Cycles: 4115

Name: RITTER2
Sterilizer ID: V2313228
11/13/2024 04:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.3 F 0.0
2:00 112.8 F 0.0
4:00 124.1 F 0.2
6:00 159.7 F 2.8
8:00 185.4 F 5.9
10:00 200.4 F 9.6
12:00 217.0 F 11.6
14:00 224.7 F 7.8
16:00 235.7 F 12.1
18:00 234.8 F 9.2
20:00 240.3 F 10.6
22:00 245.9 F 13.6
24:00 257.3 F 19.7
26:00 267.5 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.2 F 29.5
2:00 271.3 F 28.8
3:00 272.1 F 29.7
4:00 273.3 F 30.2

Min 271.0 F 28.8
Max 273.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 205.9 F 0.0
10:00 213.2 F 0.0
15:00 207.6 F 0.0
20:00 194.8 F 0.0
25:00 193.2 F 0.0
30:00 196.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 27:28
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:06:16
CYCLE COMPLETE

11/13/2024 05:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____