
Midmark M11 - v1.0.5
Total Cycles: 3353

Name: RITTER2
Sterilizer ID: V2313228
02/05/2024 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0

C383: HEATUP MODE
DOOR OPEN
CHAMBER TEMP 69F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

02/05/2024 09:22 AM