
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005908
14:54 05/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:38
STERILIZING 3:27
136.0 C 220kPa 3:27
135.9 C 221kPa 3:57
136.4 C 224kPa 4:27
136.4 C 223kPa 4:57
136.2 C 220kPa 5:27
136.0 C 221kPa 5:57
136.2 C 221kPa 6:27
136.4 C 225kPa 6:57
136.4 C 225kPa 6:57
VENTING 6:57
AIR DRYING 7:48
CYCLE COMPLETE 7:51

Digital Signature #
FC159B5A08C42C8D
