
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002357
11:34 05/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:05
STERILIZING 3:11
136.5 C 226kPa 3:11
136.3 C 224kPa 3:41
136.4 C 223kPa 4:11
136.4 C 225kPa 4:41
136.4 C 225kPa 5:11
136.4 C 225kPa 5:41
136.4 C 225kPa 6:11
136.3 C 224kPa 6:41
136.3 C 224kPa 6:41
VENTING 6:41
AIR DRYING 7:33
CYCLE COMPLETE 7:40

Digital Signature #
4BE6D414BC389C7F
