
Midmark M11 - v1.0.5
Total Cycles: 499

Name: RITTER2
Sterilizer ID: V2313228
10/28/2021 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.7 F 0.0
2:00 177.8 F 3.6
4:00 198.4 F 9.3
6:00 219.5 F 6.9
8:00 232.5 F 10.8
10:00 234.2 F 8.9
12:00 239.0 F 9.9
14:00 249.9 F 15.5
16:00 262.5 F 22.8
18:00 271.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.1 F 29.1
2:00 272.4 F 29.7
3:00 271.7 F 28.9
4:00 272.5 F 29.7

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 213.2 F 0.0
10:00 223.6 F 0.0
15:00 219.9 F 0.0
20:00 217.1 F 0.0
25:00 215.4 F 0.0
30:00 212.5 F 0.0

Drying Complete

FILLING: 2:50
HEATING: 18:08
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

10/28/2021 02:19 PM
ACCEPTED: X
REJECTED:
Operator: U02