
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000063
11:11 03/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:11
STERILIZING 4:14
136.0 C 221kPa 4:14
135.9 C 222kPa 4:44
136.4 C 222kPa 5:14
136.4 C 225kPa 5:44
136.4 C 221kPa 6:14
136.2 C 223kPa 6:44
136.1 C 223kPa 7:14
136.1 C 223kPa 7:44
136.1 C 223kPa 7:44
VENTING 7:44
AIR DRYING 8:35
CYCLE COMPLETE 8:46

Digital Signature #
AE48D8AAA0BD55EB
