
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002458
15:33 27/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:14
STERILIZING 4:22
136.4 C 225kPa 4:22
136.2 C 224kPa 4:52
136.4 C 226kPa 5:22
136.2 C 224kPa 5:52
136.4 C 226kPa 6:22
136.4 C 225kPa 6:52
136.4 C 226kPa 7:22
136.4 C 224kPa 7:52
136.4 C 224kPa 7:52
VENTING 7:52
AIR DRYING 8:54
CYCLE COMPLETE 9:16

Digital Signature #
BA6B97C927C1B4ED
