
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000369
11:18 31/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:43
STERILIZING 5:04
136.5 C 223kPa 5:04
136.4 C 225kPa 5:34
136.1 C 223kPa 6:04
136.4 C 224kPa 6:34
136.4 C 222kPa 7:04
136.2 C 222kPa 7:34
136.3 C 223kPa 8:04
136.3 C 223kPa 8:34
136.3 C 223kPa 8:34
VENTING 8:34
AIR DRYING 9:25
CYCLE COMPLETE 9:41

Digital Signature #
063539CEF8891E35
