
Midmark M11 - v1.0.5
Total Cycles: 2170

Name: RITTER2
Sterilizer ID: V2313228
03/22/2023 02:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.7 F 0.0
2:00 191.8 F 4.1
4:00 206.3 F 9.6
6:00 222.7 F 7.4
8:00 233.6 F 11.1
10:00 234.7 F 9.0
12:00 238.5 F 9.6
14:00 249.1 F 15.1
16:00 261.6 F 22.2
18:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.4
2:00 271.4 F 28.9
3:00 272.5 F 29.5
4:00 271.8 F 29.0

Min 271.3 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 229.8 F 0.0
10:00 232.2 F 0.0
15:00 229.3 F 0.0
20:00 231.1 F 0.0
25:00 230.1 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 18:23
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

03/22/2023 03:39 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____